

Registered AD



Government of Goa
Office of the Dean
Goa Medical College

Purchase Section, Bambolim- Goa
Tel.(0832) 2495163 Fax No.(0832) 2458728

No. GMC/PS/Multipara Monitor/Oncology/2024-25/ 2619

Date: 26/06/2024


TENDER ENQUIRY

1. Tenders are invited by the Dean, Goa Medical College, Bambolim – Goa on behalf of the Governor of Goa for the supply of Multipara Monitors (02 nos.), at the Dept. of Medical Oncology (Ward no. 61), Goa Medical College, Bambolim –Goa as per Annexure “A”.

The last date for the receipt of the quotation is at **3.30 pm on 24/06/2024**, which will be opened on the same day at **4.30pm**. Tender received after that date will not be entertained. The sealed tender should be super scribed as **“Tender for Supply of Multipara Monitors (02 nos.) at Dept. of Medical Oncology (Ward no. 61), Goa Medical College, Bambolim –Goa”**.

2. The Tender documents should accompany the following documents:
 - a. Registration Certificate of the company / firm.
 - b. GST registration certificate.
 - c. Partnership deed in case of partnership firm.
 - d. Copy of PAN Card, duly certified displaying the PAN No.
3. The full specification, make and guarantee period of articles should be indicated in the tender. Installation charges if any should be quoted separately, in case of instruments the list of the standard accessories supplied with the instruments should be given, spares accessories should be quoted separately, wherever possible a description pamphlet of the instruments should be enclosed.
4. The rates should be quoted F.O.R destination, Bambolim-Goa.
5. Delivery is required immediately. If the supplier is unable to supply the stores immediately, they may furnish their earliest guarantee date of delivery.
6. The rates at which GST will be charged should be specifically mentioned separately, failing which no claim for sales tax will be entertained at a future date. The rate should be inclusive of all taxes and charges.
7. The payment towards the supply will be made only after receipt of goods and also after checking the same for quality specifications etc.
8. The tender should remain open for acceptance for one year from the date of opening tender. In case, the agency tries to raise the rates during this period and refuses to supply the items as per the quoted rates, EMD/security deposit will be forfeited to the Government.

9. An earnest money deposit is required to be sent along with the tender @ 2% of the total value quoted in the tender, if the value of the tender submitted by you amounts to over one lakh. The successful company is supposed to furnish the security deposits @ 5% of the total value of the order.
10. The goods will not be insured by this office. The supplier will be responsible for the safe delivery of the goods dispatched by him to the consignee and insure the goods at his own cost.
11. Goods dispatched by V.P.P will not be accepted.
12. The offer should be firm. Conditional offer will not be considered. The rates should be valid for a period of six months from the date of the firm order.
13. The tender should be neatly typed. Corrections and erasures will not be considered unless authenticated.
14. Dean, Goa Medical College, reserves the right to accept or reject any or reject any or all the tenders without assigning any reasons thereof.


Asst. Accounts Officer (PS)
Goa Medical College
Bambolim-Goa

I accept the above terms and conditions
SIGNAURE OF TENDERER


NAME AND ADDRESS: _____

Copy to :-

1. Professor & Head, Department of Medical Oncology, Goa Medical College, Bambolim.
- ✓ 2. IT Cell, Goa Medical College Bambolim (for publication on Goa Medical College Website)
3. Guard File
4. File.

“ANNEXURE A”

Sr.no	SPECIFICATION	Qty.
1	Specification for Multipara Monitor	
	<ul style="list-style-type: none">• Three parameter table top monitor for monitoring SPO2, ECG and NIBP with flat touch keypad.• Tilttable display with high resolution 7” color TFT LCD screen.• Should be suitable for adult and Pediatric applications.• SPO2 should be with Pulse Rate Range of 30-240 bpm & Resolution 1%.• Should have Pitch Tone variation for SPO2 Monitoring• Should have 3-lead ECG facility• Should have NIBP modes like Manual, Automatic and Turbo.• Should be able to monitor parameters like Systolic, Diastolic and Mean Arterial pressure with mmHg as measuring unit.• Should have Arrhythmia detection and ST analysis.• Should have graphical and tabular trends of minimum 24 hrs with zoom option.• Should have graded and color coded visual/audio alarms.• HR/PR source selection facility from SPO2, automatic with HR Range of 30-250 bpm• Should have inbuilt rechargeable battery with 120 min back up and indication for battery charging.• Power supply should be through DC adapter• Should have optional networking facility with central nursing station.• Should weigh less than 3 kg• The manufacturing company should be ISO certified.	02
2	System should include:	
	<ul style="list-style-type: none">• Re-usable NIBP Cuff Pediatric, Neonatal, Adult — 01 no. each.• SPO2 sensor & extension cable for adult — 01 no.• Power Chord - 1 no.• User Manual - 1 No.	


Asst. Accounts Officer (PS)
Goa Medical College,
Bambolim – Goa